



CERTIFICATE OF ACCREDITATION

ANSI National Accreditation Board

11617 Coldwater Road, Fort Wayne, IN 46845 USA

This is to certify that

**Greenwich Instrument Co. Inc.,
a division of Parker Medical Inc.**

**137 New Milford Road East
Bridgewater, CT 06752**

has been assessed by ANAB and meets the requirements of international standard

ISO/IEC 17025:2017

while demonstrating technical competence in the field of

CALIBRATION

Refer to the accompanying Scope of Accreditation for information regarding the types of activities to which this accreditation applies

AC-1406

Certificate Number


ANAB Approval

Certificate Valid Through: 08/20/2021
Version No. 005 Issued: 08/12/2019



This laboratory is accredited in accordance with the recognized International Standard ISO/IEC 17025:2017.
This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system (refer to joint ISO-ILAC-IAF Communiqué dated April 2017).

SCOPE OF ACCREDITATION TO ISO/IEC 17025:2017

Greenwich Instrument Co. Inc., a division of Parker Medical Inc.

137 New Milford Road East
Bridgewater, CT 06752
Timothy Holland 860-350-4304
t.holland@parkermed.com

CALIBRATION

Valid to: **August 20, 2021**

Certificate Number: **AC-1406**

Electrical – DC/Low Frequency

Parameter / Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method and/or Equipment
Dynalizers and High Voltage Dividers DC High Voltage (kVp)	Nominal Voltage Ratio: 1 k, 10 k, 100 k (5 to 30) kV	0.22 % of Voltage Ratio	Spellman HVD-200 Voltage Divider, Spellman SL 300 Voltmeter
Dynalizers and High Voltage Dividers DC High Voltage (kVp)	Nominal Voltage Ratio: 1 k, 10 k, 100 k (30 to 150) kV	0.2 % of Voltage Ratio	Spellman HVD-200 Voltage Divider, Fluke 8845A Multimeter
Dynalizers and High Voltage Dividers Voltage Divider Frequency Response Anode Current Sensor	Nominal Voltage Ratio: 1 k, 10 k, 100 k DC to 500 Hz 500 Hz to 30 kHz (1 to 500) mA	0.6 % of Voltage Ratio 2 % of Voltage Ratio 0.1 % of Voltage Ratio	Fluke 8845A Multimeter
Dynalizers and High Voltage Dividers Filament Current Sensor	Nominal Voltage Ratio: 1 k, 10 k, 100 k (1 to 8) A	0.3 % of Voltage Ratio	Fluke 8842A Multimeter, Fluke 80J-10 shunt
Dynalyzer Digital Displays Peak Voltage Anode Current Filament Current	(20 to 150) kV 1 mA to 1 A (1 to 10) A	0.1 % of reading 0.1 % of reading 0.2 % of reading	Fluke 8845A Multimeter
Dynalyzer Digital Displays Exposure time	50 ms to 1.5 s	1 ms ± 0.01 % of reading	HP5316B Counter

Electrical – DC/Low Frequency

Parameter / Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method and/or Equipment
MAS Meter	MAS MA	0.05 % of reading 0.05 % of reading	Fluke 8845A Multimeter HP5316B Counter
DC Voltage - Source	Up to 330 mV 330 mV to 3.3 V (3.3 to 33) V (33 to 330) V 330 V to 1.02 kV	60 $\mu\text{V/V} + 3 \mu\text{V}$ 50 $\mu\text{V/V} + 5 \mu\text{V}$ 50 $\mu\text{V/V} + 50 \mu\text{V}$ 55 $\mu\text{V/V} + 500 \mu\text{V}$ 55 $\mu\text{V/V} + 1.5 \text{ mV}$	Fluke 5500A SC600 Multiproduct Calibrator
Resistance - Source	Up to 11 Ω (11 to 33) Ω (33 to 110) Ω (110 to 330) Ω 330 Ω to 1.1 k Ω (1.1 to 3.3) k Ω (3.3 to 11) k Ω (11 to 33) k Ω (33 to 110) k Ω (110 to 330) k Ω 330 k Ω to 1.1 M Ω (1.1 to 3.3) M Ω (3.3 to 11) M Ω (11 to 33) M Ω (33 to 110) M Ω (110 to 330) M Ω	120 $\mu\Omega/\Omega + 8 \text{ m}\Omega$ 120 $\mu\Omega/\Omega + 15 \text{ m}\Omega$ 90 $\mu\Omega/\Omega + 15 \text{ m}\Omega$ 90 $\mu\Omega/\Omega + 15 \text{ m}\Omega$ 90 $\mu\Omega/\Omega + 60 \text{ m}\Omega$ 90 $\mu\Omega/\Omega + 60 \text{ m}\Omega$ 90 $\mu\Omega/\Omega + 600 \text{ m}\Omega$ 90 $\mu\Omega/\Omega + 600 \text{ m}\Omega$ 110 $\mu\Omega/\Omega + 6 \Omega$ 120 $\mu\Omega/\Omega + 6 \Omega$ 150 $\mu\Omega/\Omega + 55 \Omega$ 150 $\mu\Omega/\Omega + 55 \Omega$ 600 $\mu\Omega/\Omega + 550 \Omega$ 1 m $\Omega/\Omega + 550 \Omega$ 5 m $\Omega/\Omega + 5.5 \text{ k}\Omega$ 5 m $\Omega/\Omega + 16.5 \text{ k}\Omega$	Fluke 5500A SC600 Multiproduct Calibrator
DC Current-Source	Up to 3.3 mA (3.3 to 33) mA (33 to 330) mA 330 mA to 2.2 A (2.2 to 11) A	130 $\mu\text{A/A} + 50 \text{ nA}$ 100 $\mu\text{A/A} + 250 \text{ nA}$ 100 $\mu\text{A/A} + 3.3 \mu\text{A}$ 300 $\mu\text{A/A} + 44 \mu\text{A}$ 600 $\mu\text{A/A} + 330 \mu\text{A}$	Fluke 5500A SC600 Multiproduct Calibrator

Electrical – DC/Low Frequency

Parameter / Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method and/or Equipment
AC Voltage-Source	(1 to 33) mV		Fluke 5500A SC600 Multiproduct Calibrator
	(10 to 45) Hz	3.5 mV/V + 20 μ V	
	45 Hz to 10 kHz	1.5 mV/V + 20 μ V	
	(10 to 20) kHz	2 mV/V + 20 μ V	
	(20 to 50) kHz	2.5 mV/V + 20 μ V	
	(50 to 100) kHz	3.5 mV/V + 33 μ V	
	(100 to 500) kHz	10 mV/V + 60 μ V	
	(33 to 330) mV		
	(10 to 45) Hz	2.5 mV/V + 50 μ V	
	45 Hz to 10 kHz	500 μ V/V + 20 μ V	
	(10 to 20) kHz	1 mV/V + 20 μ V	
	(20 to 50) kHz	1.6 mV/V + 40 μ V	
	(50 to 100) kHz	2.4 mV/V + 170 μ V	
	(100 to 500) kHz	37 mV/V + 330 μ V	
	330 mV to 3.3 V		
	(10 to 45) Hz	1.5 mV/V + 250 μ V	
	45 Hz to 10 kHz	300 μ V/V + 60 μ V	
	(10 to 20) kHz	800 μ V/V + 60 μ V	
	(20 to 50) kHz	1.4 mV/V + 300 μ V	
	(50 to 100) kHz	2.4 mV/V + 1.7 mV	
	(100 to 500) kHz	5 mV/V + 3.3 mV	
	(3.3 to 33) V		
	(10 to 45) Hz	1.5 mV/V + 2.5 mV	
	45 Hz to 10 kHz	400 μ V/V + 600 μ V	
	(10 to 20) kHz	800 μ V/V + 2.6 mV	
	(20 to 50) kHz	1.9 mV/V + 5 mV	
	(50 to 100) kHz	2.4 mV/V + 17 mV	
	(33 to 330) V		
	45 Hz to 1 kHz	500 μ V/V + 6.6 mV	
	(1 to 10) kHz	800 μ V/V + 15 mV	
	(10 to 20) kHz	900 μ V/V + 33 mV	
	330 V to 1.02 kV		
	45 Hz to 1 kHz	500 μ V/V + 80 mV	
	(1 to 5) kHz	2 mV/V + 100 mV	
	(5 to 10) kHz	2 mV/V + 500 mV	

Electrical – DC/Low Frequency

Parameter / Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method and/or Equipment
AC Current-Source	(29 to 330) μ A		Fluke 5500A SC600 Multiproduct Calibrator
	(10 to 20) Hz	2.5 mA/A + 150 nA	
	(20 to 45) Hz	1.25 mA/A + 150 nA	
	45 Hz to 1 kHz	1.25 mA/A + 250 nA	
	(1 to 5) kHz	4 mA/A + 150 nA	
	(5 to 10) kHz	12.5 mA/A + 150 nA	
	330 μ A to 3.3 mA		
	(10 to 20) Hz	2 mA/A + 300 nA	
	(20 to 45) Hz	1 mA/A + 300 nA	
	45 Hz to 1 kHz	1 mA/A + 300 nA	
	(1 to 5) kHz	2 mA/A + 300 nA	
	(5 to 10) kHz	6 mA/A + 300 nA	
	(3.3 to 33) mA		
	(10 to 20) Hz	2 mA/A + 3 μ A	
	(20 to 45) Hz	1 mA/A + 3 μ A	
	45 Hz to 1 kHz	900 μ A/A + 3 μ A	
	(1 to 5) kHz	2 mA/A + 3 μ A	
	(5 to 10) kHz	6 mA/A + 3 μ A	
	(33 to 330) mA		
	(10 to 20) Hz	2 mA/A + 30 μ A	
	(20 to 45) Hz	1 mA/A + 30 μ A	
	45 Hz to 1 kHz	900 μ A/A + 30 μ A	
	(1 to 5) kHz	2 mA/A + 30 μ A	
	(5 to 10) kHz	6 mA/A + 30 μ A	
	330 mA to 2.2 A		
	(10 to 45) Hz	2 mA/A + 300 μ A	
	45 Hz to 1 kHz	1 mA/A + 300 μ A	
	(1 to 5) kHz	7.5 mA/A + 300 μ A	
	(2.2 to 11) A		
	(45 to 65) Hz	600 μ A/A + 2 mA	
	(65 to 500) Hz	1 mA/A + 2 mA	
	500 Hz to 1 kHz	3.3 mA/A + 2 mA	

Calibration and Measurement Capability (CMC) is expressed in terms of the measurement parameter, measurement range, expanded uncertainty of measurement and reference standard, method, and/or equipment. The expanded uncertainty of measurement is expressed as the standard uncertainty of the measurement multiplied by a coverage factor of 2 ($k=2$), corresponding to a confidence level of approximately 95%.

Notes:

1. This scope is formatted as part of a single document including Certificate of Accreditation No. AC-1406.



Vice President